



## EC declaration of conformity Soft Tilt by GDV Technology A/S

GDV Technology A/S, Værkstedsgården 15, st. th., 2620 Albertslund, Denmark, hereby certify that the product meets the essential requirements defined in the Medical Device Directive standards:

|                          |   |
|--------------------------|---|
| DS / EN 12182: 2012      | Assistive products for persons with disabilities - General requirements and test methods.   |
| DS / EN 60601-1: 2005    | Medical electrical equipment. Part 1: General safety and essential functional requirements.<br>(Incl. Corr. 1: 2008, Corr, 2: 2008, AC: 2013 A1: 2013)<br>All parts of the standard exclusive parts excepted / changed in DS / EN 60601-2-52: 2010.   |
| DS / EN 60601-2-52: 2010 | Medical electrical equipment. Part 2-52: Particular requirements for basic safety and essential performance of hospital and care beds for medical use.<br>(Incl. AC: 2011 and A1: 2015)<br>Points: 201.8 Electrical hazards, 201.9.2.2.2 Gaps, 201.9.2.3.1 Unintended movement, 201.9.8 Support System (relevant parts), 201.11.6.5.101 Ingress of water, 201.11.6.6 Cleaning and disinfection, 201.11.8 Interruption of the power supply, 201.15.3.4.1 Hand-held ME equipment. |

We ensure and declare that:

1. The equipment satisfies current version of Council Directive 93/42 / EEC and the Danish implementation of this Directive.
2. The equipment is classified in Class 1.
3. Planning and preparation is done in accordance with company quality system in accordance with the provisions of the Directive and follows applicable requirements of DS / EN ISO 13485/2003.

Albertslund 19'th of November 2019

Michael Kock  
CEO

GDV Technology A/S